

af



A PORTABLE CALIBRATOR FOR INTERMEDIATE RANGE VACUUM GAGES

G. D. Arney, Jr. and W. F. Henderson

ARO, Inc.

May 1967

Distribution of this document is unlimited. It may be released to the Clearinghouse, Department of Commerce for sale to the general public.

**VON KÁRMÁN GAS DYNAMICS FACILITY
ARNOLD ENGINEERING DEVELOPMENT CENTER
AIR FORCE SYSTEMS COMMAND
ARNOLD AIR FORCE STATION, TENNESSEE**

NOTICES

When U. S. Government drawings, specifications, or other data are used for any purpose other than a definitely related Government procurement operation, the Government thereby incurs no responsibility nor any obligation whatsoever, and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data, is not to be regarded by implication or otherwise, or in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

Qualified users may obtain copies of this report from the Defense Documentation Center.

References to named commercial products in this report are not to be considered in any sense as an endorsement of the product by the United States Air Force or the Government.

A PORTABLE CALIBRATOR FOR
INTERMEDIATE RANGE VACUUM GAGES

G. D. Arney, Jr. and W. F. Henderson
ARO, Inc.

Distribution of this document is unlimited. It may be released to the Clearinghouse, Department of Commerce for sale to the general public.

FOREWORD

The work reported herein was done at the request of Headquarters, Arnold Engineering Development Center (AEDC), Air Force Systems Command (AFSC), under Program Element 62405334, Project 8953, Task 895306.

The results of research presented were obtained by ARO, Inc. (a subsidiary of Sverdrup & Parcel and Associates, Inc.) contract operator of AEDC, AFSC, Arnold Air Force Station, Tennessee, under Contract AF40(600)-1200. The research was conducted under ARO Project No. VL5705, and the manuscript was submitted for publication on March 22, 1967.

The authors take this opportunity to thank Mr. M. W. Bates, who did the detailed design of the calibrator.

Information in this report is embargoed under the Department of State International Traffic in Arms Regulations. This report may be released to foreign governments by departments or agencies of the U. S. Government subject to approval of the Arnold Engineering Development Center (AEDC) or higher authority within the Department of the Air Force. Private individuals or firms require a Department of State export license.

This technical report has been reviewed and is approved.

Harry R. McDaniel
2/Lt, USAF
Research Division
Directorate of Plans and Technology

Edward R. Feicht
Colonel, USAF
Director of Plans and Technology

ABSTRACT

A portable calibrator to facilitate "in-place" calibrations of intermediate range pressure transducers has been developed. The operating principle of the calibrator is to transfer gas, in controllable mass increments, into an initially evacuated vessel of known volume by use of a transfer volume. By adjusting the mass increments to specific predetermined levels, the pressure in the initially evacuated vessel can be increased in steps of 10^{-3} , 10^{-2} , 10^{-1} , or 1 torr. The pressure ranges covered are from 10^{-3} to 1 torr and from 1 to 30 torr. The indicated corresponding limits of error applicable to each range are ± 1.3 and ± 0.76 percent of desired pressure level, respectively.

CONTENTS

	<u>Page</u>
ABSTRACT.	iii
NOMENCLATURE.	v
I. INTRODUCTION	1
II. SYSTEM DESCRIPTION AND THEORY OF OPERATION	1
III. DETERMINATION OF SYSTEM VOLUME RATIOS. . . .	5
IV. DISCUSSION OF ERROR SOURCES.	6
V. CONCLUSIONS	9
REFERENCES	9

ILLUSTRATIONS

Figure

1. Calibrator Schematic and Transfer Volume Detail . .	2
2. Calibrator	3
3. Comparison of the Calibrator with an Oil Manometer .	7
4. Comparison of a Variable Capacitance Pressure Transducer with the Calibrator.	8

TABLE

I. Summary of Calibrator Error Sources.	8
---	---

NOMENCLATURE

M_{1A}	Mass of gas in V_{1A}
M_2	Mass of gas in V_2
n	Number of mass additions to V_2
ΔP	Differential pressure, $(P_3 - P_{2(n-1)})$ torr (0°C)
$P_{2(n)}$	Absolute pressure in V_2 after the n th mass addition to V_2 , torr (0°C)

$\Delta P_{2(n)}$	Incremental pressure rise in $V_2 = P_{2(n)} - P_{2(n-1)}$, torr (0°C)
$P_{2(n-1)}$	Absolute pressure in V_2 just before the nth mass addition to V_2 , torr (0°C)
P_3	Absolute pressure in V_3 , torr (0°C)
V_{1A}	Transfer volume, approximately 0.033 in. ³ (0.5408 cc)
V_{1B}	Transfer volume, approximately 0.33 in. ³ (5.408 cc)
V_{1C}	Transfer volume, approximately 3.33 in. ³ (54.08 cc)
V_2	Reference volume, 1832 in. ³ (30,021 cc)
ΔV_2	Volume internal to the gage under calibration plus the volume of the external tubulation, in. ³
V_3	N ₂ reservoir volume

SECTION I INTRODUCTION

In the VKF Aerophysics Branch there is frequent need to measure pressures covering a span from 1 atm to 2×10^{-2} torr. Precision, variable capacitance transducers with remote potentiometer readout constitute the most used pressure measuring systems. To facilitate frequent verification of system accuracy, in the range of pressures below roughly 30 torr, a portable calibrator, described herein, was designed and developed to make periodic "in-place" calibration checks of the transducers.

The operating principle of the calibrator is to transfer gas, by use of a transfer volume, in controllable mass increments, into an initially evacuated vessel of known volume. By adjusting the mass increments to predetermined levels, the pressure rise in the initially evacuated vessel can be controlled in steps of 10^{-3} , 10^{-2} , 10^{-1} , or 1 torr. Consequently, it is easy to compute the pressure existing after any number of incremental mass additions.

Other calibrators operating on the above principle have been reported (Ref. 1), and in fact at least one device is sold commercially (Ref. 2). The calibrator reported herein is somewhat different from those in Refs. 1 and 2 in that it is portable, and the pressure rise in the initially evacuated vessel is always constant for a given transfer volume. This calibrator has been in use for over 1 yr.

SECTION II SYSTEM DESCRIPTION AND THEORY OF OPERATION

The calibrator is presented schematically in Fig. 1 and pictorially in Fig. 2. Before any calibration steps are started, V_{1A} , V_{1B} , V_{1C} , V_2 , V_3 , and the volume internal to the gage under calibration are evacuated to less than 10^{-3} torr as measured with a Penning-type ionization gage connected to V_2 . Then V_{1A} , V_{1B} , V_{1C} , and V_3 are closed off from V_2 and the gage under calibration. Volume V_2 and the gage under calibration are then evacuated to less than 10^{-6} torr. Using V_2 as a reference, V_{1A} , V_{1B} , V_{1C} , and V_3 are charged with dry nitrogen to the desired differential pressure, $\Delta P = P_3 - P_{2(n-1)}$, which is determined with the mercurial barometer. The object now is to transfer, in successive, controlled increments, a mass of gas from V_3 to V_2 using




Fig. 1 Calibrator Schematic and Transfer Volume Detail




Fig. 2 Calibrator

any one of the three transfer volumes, V_{1A} , V_{1B} , or V_{1C} .^{*} Each successive mass transfer establishes a new pressure level, $P_{2(n)}$ in V_2 , which is used for calibration purposes.

Selection of the one transfer volume, of the three available transfer volumes to be vented to V_2 , is determined by the pressure steps desired in V_2 . By selecting the proper transfer volume and adjusting ΔP to the desired level, $P_{2(n)}$ can be increased in steps of 10^{-3} , 10^{-2} , 10^{-1} , and 1 torr. After each mass transfer, the transfer volume used is again vented to V_3 , and ΔP is restored to the desired level by simple manipulation of the manually controlled piston provided in V_3 and monitored using the mercurial barometer. To facilitate a very accurate measurement of ΔP , the selection of volumes used in the system design is such that the desired steps in $P_{2(n)}$ can be attained by adjusting ΔP to a value not less than 50 torr.

If ΔP is restored to the same constant level before each mass transfer, if the same transfer volume is used for each pressure increase, and if the system is isothermal, then the successive pressure rises in V_2 will be in equal steps; the proof of this follows.

Before the n th incremental mass addition to V_2 by a transfer volume, say V_{1A} , the latter having been exposed to V_3 and charged to an absolute pressure of $\Delta P + P_{2(n-1)}$, the mass of gas in V_{1A} can be expressed as

$$M_{1A(n-1)} = \frac{V_{1A} [\Delta P + P_{2(n-1)}]}{RT} \quad (1)$$

and in V_2 ,

$$M_{2(n-1)} = \frac{V_2 P_{2(n-1)}}{RT} \quad (2)$$

Now if V_{1A} is repositioned so that a passage exists between V_{1A} and V_2 , $M_{1A(n-1)}$ and $M_{2(n-1)}$ are distributed through V_{1A} and V_2 such that

$$M_{1A(n-1)} + M_{2(n-1)} = \frac{(V_{1A} + V_2) P_{2(n)}}{RT} \quad (3)$$

Substituting Eqs. (1) and (2) into (3) yields, for the isothermal case, the general equation

$$P_{2(n)} = \frac{V_{1A} \Delta P}{V_{1A} + V_2} + P_{2(n-1)} \quad (4)$$

^{*}The transfer volumes are in pneumatic communication with either V_2 or V_3 but not with both simultaneously.

Or, where $\Delta P_{2(n)} = P_{2(n)} - P_{2(n-1)}$

$$\Delta P_{2(n)} = \frac{V_{1A} \Delta P}{V_{1A} + V_2} \quad (5)$$

In practice, the calibration mode described above is only used for calibrations at pressures below 30 torr. Calibrations above this pressure are accomplished by direct comparison with the mercurial barometer.

SECTION III DETERMINATION OF SYSTEM VOLUME RATIOS

Because of the minuteness of the transfer volumes needed to produce the required volume ratios, it was not possible to fabricate these volumes to a high degree of accuracy. Hence, it was necessary to resort to experimental techniques to determine their relative sizes as compared to the computed 1832-in.³ volume of V_2 . This was accomplished by using an estimated value for ΔP and adding sufficient mass increments in V_2 such that the final $P_{2(n)}$ could be measured accurately with an oil manometer.* Dividing the final $P_{n(2)}$ by the number of mass transfer increments required to produce it yielded an average value for $\Delta P_{2(n)}$. Inserting these pressure relations into Eq. (5) produced the following:

$$(V_2 + V_{1A})/V_{1A} = 59,693.0$$

$$(V_2 + V_{1B})/V_{1B} = 5666.2$$

$$(V_2 + V_{1C})/V_{1C} = 560.0$$

This process was repeated many times with no noticeable change in the volume ratios.

In practice, the gage under calibration and its tubulation may significantly alter the volume V_2 . Hence, in establishing calibration

*The oil manometer used here is classified as a secondary standard and is calibrated against AEDC reference standards which are traceable to the National Bureau of Standards.

settings for ΔP , it may be necessary to account for this.* The method of doing so is described in the following relations which were derived from Eq. (5) and the above volume ratios:

$$\Delta P = \frac{V_1 + V_2 + \Delta V_2}{V_1} \left(\text{desired } \Delta P_{2(n)} \right) \text{ torr}$$

For transfer volume V_{1A} :

$$\Delta P_A = \left(59,693 + \frac{\Delta V_2}{0.033} \right) \left(\text{desired } \Delta P_{2(n)} \right) \text{ torr}$$

For transfer volume V_{1B} :

$$\Delta P_B = \left(5666.2 + \frac{\Delta V_2}{0.33} \right) \left(\text{desired } \Delta P_{2(n)} \right) \text{ torr}$$

For transfer volume V_{1C} :

$$\Delta P_C = \left(560.0 + \frac{\Delta V_2}{3.33} \right) \left(\text{desired } \Delta P_{2(n)} \right) \text{ torr}$$

(ΔV_2 = in.³ in these formulas.)

SECTION IV DISCUSSION OF ERROR SOURCES

The first obvious question regards the experimental techniques for determining the volume ratios between the transfer volumes and V_2 , i. e., is the mass transfer method sufficiently consistent to justify taking an average of many incremental additions. Since $\Delta P > 50$ torr and can be determined with an uncertainty no greater than ± 0.3 torr, the question reduces to determining the consistency of the transfer volumes. The only conceivable variables in this case are O-ring slippage, wear, or deformation. Simple calculations have convinced the authors that for any of the transfer volumes, there should be no more than a ± 0.5 -percent inconsistency.

Since no method of temperature control is built into the system, another question concerns the assumption of an isothermal process. This was not considered to be a problem because the calibrator is used in an environment where the ambient temperature does not fluctuate more than $\pm 3.0^\circ\text{F}$ and because the thermal mass of the calibrator is

*Though not previously mentioned, the volume of the oil manometer was accounted for in establishing the given volume relations.

sufficiently large that operation should not significantly alter its temperature. In addition, small temperature differences can be tolerated. For example, a 1.0°F variation in the temperature in V₂ leads to, roughly, a 0.2-percent variation in determining P_{2(n)}.

Other possible sources of error include condensables, surface adsorption, leaks, and uncertainty in system "zero". It was anticipated that the first two of these could be avoided by using dry nitrogen for the working gas and stainless steel for the walls of V₁, A, B, C, V₂, and V₃. Only extreme care, of course, could avoid leaks. As for the system zero, this does not pose a problem because V₂ is pumped down to an absolute pressure at least three orders of magnitude below the smallest calibration increment.

For want of better techniques for proving the adequacy of the calibrator, it was checked against the secondary standard oil manometer mentioned previously over the appropriate operating range (Fig. 3).




Fig. 3 Comparison of the Calibrator with an Oil Manometer

A later comparison of the calibrator with a variable capacitance pressure transducer which had been recently calibrated against an AEDC reference standard is presented in Fig. 4. The resolution of the variable capacitance pressure transducer is not sufficient to permit precise

pressure measurements below 10^{-1} torr. This accounts for the deviation from the calibrator pressure below 10^{-1} torr evident in Fig. 4.

$$\text{Deviation} = \frac{(\text{Capacitance Transducer Reading}) - (\text{Calibrator Pressure Level})}{\text{Calibrator Pressure Level}} \times 100 \text{ percent}$$




Fig. 4 Comparison of a Variable Capacitance Pressure Transducer with the Calibrator

An analysis of the sources and limits of error when calibrating with this system is summarized and presented in Table I.

TABLE I
SUMMARY OF CALIBRATOR ERROR SOURCES

Error Source	Error*, percent		Basis for Error Determination
	Calibration Range		
	10 ⁻³ to 1 torr	1 to 30 torr	
Fluctuation in differential pressure (ΔP)	±0.5	±0.05	Uncertainty inherent in the mercurial barometer
Uncertainty in volume ratio V ₁ /(V ₂ + ΔV ₂ + V ₁)	±0.5	±0.5	Observations over many calibration runs
Fluctuations in system temperature	±0.2	±0.2	Observations
Fluctuations in system zero	±0.1	±0.01	Uncertainty in measuring residual pressure in V ₂
Maximum error in establishing a desired P _{2(n)}	±1.3	±0.76	

*The term "error", as used here, is in percent of the desired pressure level and is the uncertainty in establishing $P_{2(n)}$.

SECTION V CONCLUSIONS

The incremental mass addition of gas to a fixed volume appears to be a suitable means of establishing pressure levels in the volume for the purpose of calibrating pressure gages. By an appropriate method, it is a simple matter to raise the pressure in known increments so that only simple addition is necessary to determine any pressure level.

No problem of condensables or adsorption seems to exist at pressure levels for which the calibrator was intended.

The calibrator is suitable for use as a portable calibration standard over the pressure range from 10^{-3} to 1 torr with the indicated capability to establish incremental pressure levels to within 1.3 percent of the desired level and over the pressure range from 1 to 30 torr with the indicated capability to establish incremental pressure levels to within 0.75 percent of the desired level.

REFERENCES

1. Smetana, F. O. and Everton, J. G. "A Calibration Facility for High Resolution Vacuum Gauges." University of Southern California, USCEC Rept. 83-214, July 1962.
2. Bottorff, M. R. and Chuan, R. L. "Traceable Vacuum-Gage Calibration by Incremental Mass Addition." Research and Development, August 1966, pp. 60-63.

UNCLASSIFIED

Security Classification

DOCUMENT CONTROL DATA - R&D

(Security classification of title, body of abstract and indexing annotation must be entered when the overall report is classified)

1 ORIGINATING ACTIVITY (Corporate author) Arnold Engineering Development Center, ARO, Inc., Operating Contractor, Arnold Air Force Station, Tennessee		2a REPORT SECURITY CLASSIFICATION UNCLASSIFIED	
		2b GROUP N/A	
3 REPORT TITLE A PORTABLE CALIBRATOR FOR INTERMEDIATE RANGE VACUUM GAGES			
4 DESCRIPTIVE NOTES (Type of report and inclusive dates) N/A			
5 AUTHOR(S) (Last name, first name, initial) Arney, G. D., Jr., and Henderson, W. F., ARO, Inc.			
6 REPORT DATE May 1967		7a TOTAL NO. OF PAGES 16	7b. NO OF REFS 2
8a CONTRACT OR GRANT NO. AF40(600)-1200		9a. ORIGINATOR'S REPORT NUMBER(S) AEDC-TR-67-83	
b. PROJECT NO. 8953			
c. Program Element 62405334		9b. OTHER REPORT NO(S) (Any other numbers that may be assigned this report) N/A	
d. Task 895306			
10 AVAILABILITY/LIMITATION NOTICES Distribution of this document is unlimited. It may be released to the Clearinghouse, Department of Commerce for sale to the general public.			
11 SUPPLEMENTARY NOTES Available in DDC.		12. SPONSORING MILITARY ACTIVITY Arnold Engineering Development Center, Air Force Systems Command, Arnold Air Force Station, Tennessee	
13 ABSTRACT A portable calibrator to facilitate "in-place" calibrations of intermediate range pressure transducers has been developed. The operating principle of the calibrator is to transfer gas, in controllable mass increments, into an initially evacuated vessel of known volume by use of a transfer volume. By adjusting the mass increments to specific predetermined levels, the pressure in the initially evacuated vessel can be increased in steps of 10^{-3} , 10^{-2} , 10^{-1} , or 1 torr. The pressure ranges covered are from 10^{-3} to 1 torr and from 1 to 30 torr. The indicated corresponding limits of error applicable to each range are ± 1.3 and ± 0.76 percent of desired pressure level, respectively.			

14 KEY WORDS	LINK A		LINK B		LINK C	
	ROLE	WT	ROLE	WT	ROLE	WT
portable calibrators vacuum gages intermediate range pressure transducers gas transfer						

INSTRUCTIONS

1. **ORIGINATING ACTIVITY:** Enter the name and address of the contractor, subcontractor, grantee, Department of Defense activity or other organization (*corporate author*) issuing the report.

2a. **REPORT SECURITY CLASSIFICATION:** Enter the overall security classification of the report. Indicate whether "Restricted Data" is included. Marking is to be in accordance with appropriate security regulations.

2b. **GROUP:** Automatic downgrading is specified in DoD Directive 5200.10 and Armed Forces Industrial Manual. Enter the group number. Also, when applicable, show that optional markings have been used for Group 3 and Group 4 as authorized.

3. **REPORT TITLE:** Enter the complete report title in all capital letters. Titles in all cases should be unclassified. If a meaningful title cannot be selected without classification, show title classification in all capitals in parenthesis immediately following the title.

4. **DESCRIPTIVE NOTES:** If appropriate, enter the type of report, e.g., interim, progress, summary, annual, or final. Give the inclusive dates when a specific reporting period is covered.

5. **AUTHOR(S):** Enter the name(s) of author(s) as shown on or in the report. Enter last name, first name, middle initial. If military, show rank and branch of service. The name of the principal author is an absolute minimum requirement.

6. **REPORT DATE:** Enter the date of the report as day, month, year, or month, year. If more than one date appears on the report, use date of publication.

7a. **TOTAL NUMBER OF PAGES:** The total page count should follow normal pagination procedures, i.e., enter the number of pages containing information.

7b. **NUMBER OF REFERENCES:** Enter the total number of references cited in the report.

8a. **CONTRACT OR GRANT NUMBER:** If appropriate, enter the applicable number of the contract or grant under which the report was written.

8b, &, & 8d. **PROJECT NUMBER:** Enter the appropriate military department identification, such as project number, subproject number, system numbers, task number, etc.

9a. **ORIGINATOR'S REPORT NUMBER(S):** Enter the official report number by which the document will be identified and controlled by the originating activity. This number must be unique to this report.

9b. **OTHER REPORT NUMBER(S):** If the report has been assigned any other report numbers (*either by the originator or by the sponsor*), also enter this number(s).

10. **AVAILABILITY/LIMITATION NOTICES:** Enter any limitations on further dissemination of the report, other than those

imposed by security classification, using standard statements such as:

- (1) "Qualified requesters may obtain copies of this report from DDC."
- (2) "Foreign announcement and dissemination of this report by DDC is not authorized."
- (3) "U. S. Government agencies may obtain copies of this report directly from DDC. Other qualified DDC users shall request through _____."
- (4) "U. S. military agencies may obtain copies of this report directly from DDC. Other qualified users shall request through _____."
- (5) "All distribution of this report is controlled. Qualified DDC users shall request through _____."

If the report has been furnished to the Office of Technical Services, Department of Commerce, for sale to the public, indicate this fact and enter the price, if known.

11. **SUPPLEMENTARY NOTES:** Use for additional explanatory notes.

12. **SPONSORING MILITARY ACTIVITY:** Enter the name of the departmental project office or laboratory sponsoring (*paying for*) the research and development. Include address.

13. **ABSTRACT:** Enter an abstract giving a brief and factual summary of the document indicative of the report, even though it may also appear elsewhere in the body of the technical report. If additional space is required, a continuation sheet shall be attached.

It is highly desirable that the abstract of classified reports be unclassified. Each paragraph of the abstract shall end with an indication of the military security classification of the information in the paragraph, represented as (TS), (S), (C), or (U).

There is no limitation on the length of the abstract. However, the suggested length is from 150 to 225 words.

14. **KEY WORDS:** Key words are technically meaningful terms or short phrases that characterize a report and may be used as index entries for cataloging the report. Key words must be selected so that no security classification is required. Identifiers, such as equipment model designation, trade name, military project code name, geographic location, may be used as key words but will be followed by an indication of technical context. The assignment of links, rules, and weights is optional.